

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS

Page 1 of 2

TO Cher Snyder, Assistant DEO
Engineering & Compliance

LABORATORY NO 1607512

DATE RECEIVED 03/16/2016

SAMPLE DESCRIBED AS
One Summa Canister, 24 h sample

FACILITY ID NO NA

REQUESTED BY Sumner Wilson

SAMPLING LOCATION
Highlands Community
12378 High Glen Way
Northridge, CA 91326


ST NO / PROJECT NA

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54618</u>
Pressure (Torr)	740
CO, ppm	< 3
CH ₄ , ppm	2
CO ₂ , ppm	393
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 3/18/16

Approved By: _____


Rudy Eden, Senior Manager
Laboratory Services
909-396-2391

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1607512</u>
One Summa Canister, 24 h sample	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC91340	MDL				
CO, ppmvC	0.3	10.40	9.90	-4.82	PASS
CH ₄ , ppmvC	0.3	10.17	10.44	2.63	PASS
CO ₂ , ppmvC	0.4	10.38	10.39	0.05	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.2	11.00	10.17	-7.55	PASS
NM/NEOC, ppmvC	0.2	10.64	10.09	-5.20	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC135067	MDL				
CO, ppmvC	0.3	10100	9648	-4.48	PASS
CH ₄ , ppmvC	0.3	9950	10015	0.65	PASS
CO ₂ , ppmvC	0.4	10100	9624	-4.71	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.2	9940	9799	-1.42	PASS
NM/NEOC, ppmvC	0.2	10000	9723	-2.77	PASS

DATE ANALYZED 03/16/2016
REFERENCE NO: 16QM2AA
 QM2-101-89

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DIS
 INV
 LA
LABC

WO #: 1607512



TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 3/16/16

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Qian Zhou Date: 3/16/16 Time: 8:40am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End vac
Highlands Community	54618	3-15-16 / 00:00 / 24 hours	-30"	+1

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhouqian</i>	<i>JL</i>	SCAQMD Lab	3/16/16	11:49

Remarks Daily scheduled samples from trailer

Highlands community pool parking lot. Address: 12378 High Glen Way, Northridge CA 91326 (across from 12377)