

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT  
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS  
REPORT OF LABORATORY ANALYSIS

Page 1 of 2


<b>TO</b>	Jason Low Atmospheric Measurements Manager Science and Technology Advancement	<b>LABORATORY NO</b>	<u>1618131</u>
<b>SAMPLE DESCRIBED AS</b>	24 hour sample Canister 54213	<b>DATE RECEIVED</b>	<u>06/30/2016</u>
<b>SAMPLING LOCATION</b>	Reseda Station 18328 Gault St Los Angeles, CA 91335	<b>FACILITY ID NO</b>	<u>NA</u>
		<b>REQUESTED BY</b>	<u>Sumner Wilson</u>
		<b>ST NO / PROJECT</b>	<u>NA</u>

---

**Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).**

Type	Canister
Number	<u>54213</u>
Pressure (Torr)	760
CO, ppm	< 3
CH <sub>4</sub> , ppm	2
CO <sub>2</sub> , ppm	434
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 7/5/16

Approved By:   
Solomon Teffera, Acting Senior Manager  
Laboratory Services  
909-396-2199

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT**  
**21865 Copley Drive, Diamond Bar CA 91765-1482**

**MONITORING & ANALYSIS**  
**QUALITY CONTROL SUMMARY**

Page 2 of 2

<b>SAMPLE DESCRIBED AS</b>	<b>LABORATORY NO</b>	1618131
24 hour sample Canister 54213	<b>REQUESTED BY</b>	Sumner Wilson

**Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).**

**QUALITY CONTROL -- End of run control recovery**

	MDL	Theoretical	Measured	Percent Difference	QC Limit	±5% or ± 1
CC106783	0.4	1.92	2.29	19.05	PASS	
CO, ppmvC	0.3	2.02	1.88	-7.24	PASS	
CH <sub>4</sub> , ppmvC	0.4	1.57	2.20	40.39	PASS	
CO <sub>2</sub> , ppmvC	0.4	NA	NA	NA	NA	
C <sub>2</sub> H <sub>4</sub> , ppmvC	0.4	2.03	1.91	-6.00	PASS	
C <sub>2</sub> H <sub>6</sub> , ppmvC	0.5	2.03	2.14	5.75	PASS	
NM/NEOC, ppmvC						

	MDL	Theoretical	Measured	Percent Difference	QC Limit	±5% or ± 1
CC12628	0.4	1036	1047	1.09	PASS	
CO, ppmvC	0.3	1068	1049	-1.80	PASS	
CH <sub>4</sub> , ppmvC	0.4	1022	1033	1.11	PASS	
CO <sub>2</sub> , ppmvC	0.4	NA	NA	NA	NA	
C <sub>2</sub> H <sub>4</sub> , ppmvC	0.4	1044	1032	-1.15	PASS	
C <sub>2</sub> H <sub>6</sub> , ppmvC	0.5	1024	1017	-0.65	PASS	
NM/NEOC, ppmvC						

<b>DATE ANALYZED</b>	06/30/2016
<b>REFERENCE NO:</b>	16QM2AD
	QM2-101-113

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT  
SAMPLE ANALYSIS REQUEST**

LA

WO #: 1618131



TO: SCAQMD LAB:  OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. \_\_\_\_\_

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: \_\_\_\_\_ City: \_\_\_\_\_ Zip: 91326

Contact Person: \_\_\_\_\_ Title: \_\_\_\_\_ Tel: \_\_\_\_\_

Analysis Requested by: Sumner Wilson Date: 6/30/16

Approved by: Jason Low Office: \_\_\_\_\_ Budget #: 44716

REASON REQUESTED: Court/Hearing Board  Permit Pending  Hazardous/Toxic Spill

Suspected Violation Rule(s) \_\_\_\_\_ Other

Sample Collected by: Qian Zhou Date: 6/30/16 Time: 11:00pm

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Reseda Station	54213	6/29/16 / 00:00 / 24 hours	<-30"	+11.5

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhouqian</i>	<i>Ar J. Liu</i>	SCAQMD Lab	<i>6/30/16</i>	<i>12:44</i>

Remarks: 1:3 scheduled samples from station at Reseda.  
Reseda Station - 18328 Gault St, Los Angeles, CA 91335  
GPS (34.199225, -118.532743). Left sampler sn 4148