

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS

Page 1 of 2


TO	Jason Low Atmospheric Measurements Manager Science and Technology Advancement	LABORATORY NO	<u>1619317</u>
SAMPLE DESCRIBED AS	24 hour sample Canister 10988	DATE RECEIVED	<u>07/12/2016</u>
SAMPLING LOCATION	Porter Ranch Community School	FACILITY ID NO	<u>NA</u>
		REQUESTED BY	<u>Sumner Wilson</u>
		ST NO / PROJECT	<u>NA</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>10988</u>
Pressure (Torr)	760
CO, ppm	< 3
CH ₄ , ppm	2
CO ₂ , ppm	425
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 7/19/16

Approved By: _____


Solomon Teffera, Acting Senior Manager
Laboratory Services
909-396-2199

7/19/16

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1619317</u>
24 hour sample Canister 10988	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.4	1.92	2.55	32.57	PASS
CH ₄ , ppmvC	0.3	2.02	2.02	-0.40	PASS
CO ₂ , ppmvC	0.4	1.57	3.04	93.58	FAIL
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	1.94	-4.48	PASS
NM/NEOC, ppmvC	0.5	2.03	2.05	0.85	PASS


		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.4	1036	1056	1.90	PASS
CH ₄ , ppmvC	0.3	1068	1051	-1.59	PASS
CO ₂ , ppmvC	0.4	1022	1037	1.45	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1032	-1.15	PASS
NM/NEOC, ppmvC	0.5	1024	1015	-0.88	PASS

DATE ANALYZED	07/14/2016
REFERENCE NO:	16QM2AD
	QM2-101-116

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DIS
 INV
 LA
LABO

WO #: 1619317



TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 7/12/16

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Qian Zhou Date: 7/12/16 Time: 11:15am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Porter Ranch Community Elementary School (PRCS)	10988	7/11/16 / 00:00 / 24 hours	<-30"	+13

Relinquished by	Received by	Firm/Agency	Date	Time
<i>zhongqian</i>	<i>Sumner Wilson</i>	SCAQMD Lab	7/12/16	15:17

Remarks: 1:3 scheduled samples from trailer at PRCS
Porter Ranch Community School Elementary – 12450 Mason Ave, Porter Ranch, CA 91326
GPS (34.293369, -118.580505)
Left sampler, sn 4083