

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS


Page 1 of 2

TO	Jason Low Atmospheric Measurements Manager Science and Technology Advancement	LABORATORY NO	<u>1619931</u>
SAMPLE DESCRIBED AS	24 hour sample Canister 54060	DATE RECEIVED	<u>07/18/2016</u>
SAMPLING LOCATION	Porter Ranch Community School	FACILITY ID NO	<u>NA</u>
		REQUESTED BY	<u>Sumner Wilson</u>
		ST NO / PROJECT	<u>NA</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54060</u>
Pressure (Torr)	760
CO, ppm	7
CH ₄ , ppm	2
CO ₂ , ppm	416
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 7/21/16

Approved By: 

Solomon Teffera, Acting Senior Manager
Laboratory Services
909-396-2199

Sumner Wilson
7/21/16

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1619931</u>
24 hour sample Canister 54060	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.4	1.92	1.87	-2.69	PASS
CH ₄ , ppmvC	0.3	2.02	1.87	-7.82	PASS
CO ₂ , ppmvC	0.4	1.57	2.31	47.51	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	1.89	-6.74	PASS
NM/NEOC, ppmvC	0.5	2.03	1.97	-2.92	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.4	1036	1059	2.23	PASS
CH ₄ , ppmvC	0.3	1068	1056	-1.15	PASS
CO ₂ , ppmvC	0.4	1022	1039	1.71	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1041	-0.26	PASS
NM/NEOC, ppmvC	0.5	1024	1019	-0.42	PASS

DATE ANALYZED	07/19/2016
REFERENCE NO:	16QM2AD
	QM2-101-116

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DISTRICT INFORMATION
 INVOICE SOURCE
 LAP AUDIT
 LABORATORY 1619931

TO: SCAQMD LAB: OTHER: _____
 SOURCE NAME: Southern California Gas Co. I.D. No. _____
 Source Address: 12801 Tampa Ave City: Porter Ranch
 Mailing Address: _____ City: _____ Zip: 91326
 Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 7/18/16
 Approved by: Jason Low Office: _____ Budget #: 44716
 REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill
 Suspected Violation Rule(s) _____ Other _____

Sample Collected by: Qian Zhou Date: 7/18/16 Time: 10:50am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Porter Ranch Community Elementary School (PRCS)	54060	7/17/16 / 00:00 / 24 hours	<-30"	+12

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhou Qian</i>	<i>[Signature]</i>	SCAQMD Lab	7/18/16	1:18 pm

Remarks: 1:3 scheduled samples from trailer at PRCS
 Porter Ranch Community School Elementary – 12450 Mason Ave, Porter Ranch, CA 91326
 GPS (34.293369, -118.580505)
 Left sampler, sn 4083