

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS


Page 1 of 2

TO	Jason Low Atmospheric Measurements Manager Science and Technology Advancement	LABORATORY NO	<u>1621115</u>
SAMPLE DESCRIBED AS	24 hour sample Canister 22487	DATE RECEIVED	<u>08/02/2016</u>
SAMPLING LOCATION	Castlebay Elementary School	FACILITY ID NO	<u>NA</u>
		REQUESTED BY	<u>Sumner Wilson</u>
		ST NO / PROJECT	<u>NA</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>22487</u>
Pressure (Torr)	760
CO, ppm	< 3
CH ₄ , ppm	3
CO ₂ , ppm	435
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 8/5/16

Approved By: 
Solomon Teffera, Acting Senior Manager
Laboratory Services
909-396-2199

Handwritten note: 967
8-3-16

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1621115</u>
24 hour sample Canister 22487	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.4	1.92	2.07	7.54	PASS
CH ₄ , ppmvC	0.3	2.02	1.87	-7.51	PASS
CO ₂ , ppmvC	0.4	1.57	7.76	394.52	FAIL
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	1.91	-6.07	PASS
NM/NEOC, ppmvC	0.5	2.03	1.82	-10.32	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.4	1036	1044	0.75	PASS
CH ₄ , ppmvC	0.3	1068	1041	-2.55	PASS
CO ₂ , ppmvC	0.4	1022	1037	1.48	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1028	-1.49	PASS
NM/NEOC, ppmvC	0.5	1024	1012	-1.11	PASS

DATE ANALYZED	08/02/2016
REFERENCE NO:	16QM2AD QM2-101-118

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DIST
 INVC
 LAP
LABOR

WO #: 1621115



TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 8/2/16

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Qian Zhou Date: 8/2/16 Time: 12:00pm

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Porter Ranch / Castlebay Elem	22487	7/29/16 / 00:00 / 24 hours	-30"	+18

Relinquished by	Received by	Firm/Agency	Date	Time
<i>zhangjian</i>	<i>Qian Zhou</i>	SCAQMD Lab	08/02/16	15:27

Remarks: 1:6 scheduled samples from trailer at Castlebay
 Castlebay Lane Charter School – 19010 Castlebay Ln, Porter Ranch, CA
 Right sampler, sn 4671