

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS


Page 1 of 2

TO	Jason Low Atmospheric Measurements Manager Science and Technology Advancement	LABORATORY NO	<u>1622944</u>
SAMPLE DESCRIBED AS	Triggered Sample Canister 54037	DATE RECEIVED	<u>08/16/2016</u>
SAMPLING LOCATION	Highlands Community Center	FACILITY ID NO	<u>NA</u>
		REQUESTED BY	<u>Sumner Wilson</u>
		ST NO / PROJECT	<u>NA</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54037</u>
Pressure (Torr)	760
CO, ppm	< 3
CH ₄ , ppm	3
CO ₂ , ppm	433
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	258

Date Approved: 08/19/16

Approved By: 
Solomon Teffera, Acting Senior Manager
Laboratory Services
909-396-2199

2744
6-19-16

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1622944</u>
Triggered Sample Canister 54037	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.4	1.92	1.70	-11.82	PASS
CH ₄ , ppmvC	0.3	2.02	1.89	-6.62	PASS
CO ₂ , ppmvC	0.4	1.57	2.19	39.46	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	1.92	-5.49	PASS
NM/NEOC, ppmvC	0.5	2.03	1.99	-1.91	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.4	1036	1038	0.15	PASS
CH ₄ , ppmvC	0.3	1068	1032	-3.35	PASS
CO ₂ , ppmvC	0.4	1022	1022	-0.04	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1023	-2.02	PASS
NM/NEOC, ppmvC	0.5	1024	994	-2.89	PASS

DATE ANALYZED	08/17/2016
REFERENCE NO:	16QM2AD
	QM2-101-119

1622944



**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DIS
 INV
 LAJ
 LABO

TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 8/16/16

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Qian Zhou Date: 8/16/16 Time: 11:50

REQUESTED ANALYSIS: PAMS anal

City/Location	Can#	Start day / time/ durat	rt vac	End vac
Highlands Community	54037	8/16/16 / 06:33 / 5 min	-30"	0

Relinquished by	Received by	Firm/Agency	Date	Time
<i>zhongqian</i>	<i>[Signature]</i>	SCAQMD Lab	8/16/2016	14:25

Remarks: Samples collected by passive sampling via the XonTech 912 triggered by the Thermo 55i NMHC.

Trigger is set to 5ppm

Highlands community pool parking lot. Address: 12378 High Glen Way, Northridge CA 91326 (across from 12377)