

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS


Page 1 of 2

TO Jason Low Atmospheric Measurements Manager Science and Technology Advancement	LABORATORY NO <u>1623627</u>
SAMPLE DESCRIBED AS 24 hour sample Canister E5753	DATE RECEIVED <u>08/23/2016</u>
SAMPLING LOCATION Highlands Community Center	FACILITY ID NO <u>NA</u>
	REQUESTED BY <u>Sumner Wilson</u>
	ST NO / PROJECT <u>NA</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>E5753</u>
Pressure (Torr)	760
CO, ppm	7
CH ₄ , ppm	3
CO ₂ , ppm	415
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 9/7/16

Approved By: 
Solomon Teffera, Acting Senior Manager
Laboratory Services
909-396-2199

141 9-6-16

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1623627</u>
24 hour sample Canister E5753	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.4	1.92	2.08	8.32	PASS
CH ₄ , ppmvC	0.3	2.02	1.91	-5.59	PASS
CO ₂ , ppmvC	0.4	1.57	12.33	685.69	FAIL
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	1.92	-5.19	PASS
NM/NEOC, ppmvC	0.5	2.03	2.23	9.72	PASS


		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.4	1036	1018	-1.69	PASS
CH ₄ , ppmvC	0.3	1068	1013	-5.15	FAIL
CO ₂ , ppmvC	0.4	1022	1005	-1.68	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1001	-4.15	PASS
NM/NEOC, ppmvC	0.5	1024	981	-4.19	PASS

DATE ANALYZED	08/23/2016
REFERENCE NO:	16QM2AD
	QM2-101-120

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DISTRICT
 INVENTORY
 LABORATORY

1623627



TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 8/23/16

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Qian Zhou Date: 8/23/16 Time: 12:00pm

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End vac
Highlands Community	E5753	8-19-16 / 00:00 / 24 hours	-30"	12
*Makeup run for 8/16/16				

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhouqian</i>	<i>[Signature]</i>	SCAQMD Lab	<i>8/23/2016</i>	<i>15:15</i>

Remarks 1:6 scheduled samples from trailer

Highlands community pool parking lot. Address: 12378 High Glen Way, Northridge CA 91326 (across from 12377)

Left sampler, sn 5624