

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT  
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS  
REPORT OF LABORATORY ANALYSIS

Page 1 of 2


<b>TO</b>	Jason Low ADEO Science and Technology Advancement	<b>LABORATORY NO</b>	<u>1628324</u>
<b>SAMPLE DESCRIBED AS</b>	24 hour sample Canister 54319	<b>DATE RECEIVED</b>	<u>10/18/2016</u>
<b>SAMPLING LOCATION</b>	Highlands Community Center	<b>FACILITY ID NO</b>	<u>NA</u>
		<b>REQUESTED BY</b>	<u>Sumner Wilson</u>
		<b>ST NO / PROJECT</b>	<u>Southern California Gas C</u>

---

**Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).**

Type	Canister
Number	<u>54319</u>
Pressure (Torr)	760
CO, ppm	5
CH <sub>4</sub> , ppm	2
CO <sub>2</sub> , ppm	418
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 10/25/16

Approved By:   
Solomon Teffera, Acting Senior Manager  
Laboratory Services  
909-396-2199

*Handwritten:* 7/24/16  
10/21/16

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT**  
**21865 Copley Drive, Diamond Bar CA 91765-1482**

**MONITORING & ANALYSIS**  
**QUALITY CONTROL SUMMARY**

Page 2 of 2

<b>SAMPLE DESCRIBED AS</b>	<b>LABORATORY NO</b>	<u>1628324</u>
24 hour sample Canister 54319	<b>REQUESTED BY</b>	<u>Sumner Wilson</u>

**Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).**

**QUALITY CONTROL -- End of run control recovery**

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.4	1.92	2.22	15.29	PASS
CH <sub>4</sub> , ppmvC	0.3	2.02	1.98	-2.09	PASS
CO <sub>2</sub> , ppmvC	0.4	1.57	5.95	279.29	FAIL
C <sub>2</sub> H <sub>4</sub> , ppmvC	0.4	NA	NA	NA	NA
C <sub>2</sub> H <sub>6</sub> , ppmvC	0.4	2.03	2.17	6.91	PASS
NM/NEOC, ppmvC	0.5	2.03	2.07	1.85	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.4	1036	1045	0.91	PASS
CH <sub>4</sub> , ppmvC	0.3	1068	1042	-2.42	PASS
CO <sub>2</sub> , ppmvC	0.4	1022	1029	0.73	PASS
C <sub>2</sub> H <sub>4</sub> , ppmvC	0.4	NA	NA	NA	NA
C <sub>2</sub> H <sub>6</sub> , ppmvC	0.4	1044	1031	-1.28	PASS
NM/NEOC, ppmvC	0.5	1024	1005	-1.79	PASS

<b>DATE ANALYZED</b>	10/18/2016
<b>REFERENCE NO:</b>	16QM2AD
	QM2-101-128

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT  
SAMPLE ANALYSIS REQUEST**

DIST  
 INVC  
 LAP  
**LABOR**

WO #: 1628324



TO: SCAQMD LAB:  OTHER:   
 SOURCE NAME: Southern California Gas Co. I.D. No. \_\_\_\_\_  
 Source Address: 12801 Tampa Ave City: Porter Ranch  
 Mailing Address: \_\_\_\_\_ City: \_\_\_\_\_ Zip: 91326  
 Contact Person: \_\_\_\_\_ Title: \_\_\_\_\_ Tel: \_\_\_\_\_

Analysis Requested by: Sumner Wilson Date: 10/18/16  
 Approved by: Jason Low Office: \_\_\_\_\_ Budget #: 44716  
 REASON REQUESTED: Court/Hearing Board  Permit Pending  Hazardous/Toxic Spill   
 Suspected Violation Rule(s) \_\_\_\_\_ Other

Sample Collected by: Bo Vongphachanh Date: 10/18/16 Time: 10:15am

**REQUESTED ANALYSIS: PAMS analysis**

City/Location	Can#	Start day / time/ duration	Start vac	End vac
Highlands Community	54319	10/15/16 / 00:00 / 24 hours	-30"	11

Relinquished by	Received by	Firm/Agency	Date	Time
		SCAQMD Lab	10/18/16	14:38

Remarks 1:3 scheduled samples from trailer  
 Highlands community pool parking lot. Address: 12378 High Glen Way, Northridge CA 91326 (across from 12377)  
 Right Sampler, sn 5323