

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS


Page 1 of 2

TO	Jason Low, Ph.D., ADEO Monitoring & Analysis	LABORATORY NO	<u>1631317</u>
SAMPLE DESCRIBED AS	24 hour sample Canister 54214	DATE RECEIVED	<u>11/10/2016</u>
SAMPLING LOCATION	Reseda Station 18328 Gault St Los Angeles, CA 91335	FACILITY ID NO	<u>NA</u>
		REQUESTED BY	<u>Sumner Wilson</u>
		ST NO / PROJECT	<u>NA</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).


Type	Canister
Number	<u>54214</u>
Pressure (Torr)	760
CO, ppm	5
CH ₄ , ppm	3
CO ₂ , ppm	496
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	2

Date Approved: 12/14/2016

Approved By: 

Aaron Katzenstein, Ph.D., Sr. Manager
Laboratory Services
909-396-2219

94
12816



SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1631317</u>
24 hour sample Canister 54214	REQUESTED BY	<u>Sumner Wilson</u>


Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

			Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL					
CO, ppmvC	0.4		1.92	1.92	0.00	PASS
CH ₄ , ppmvC	0.3		2.02	2.18	7.56	PASS
CO ₂ , ppmvC	0.4		1.57	7.43	373.34	FAIL
C ₂ H ₄ , ppmvC	0.4		NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4		2.03	2.05	0.83	PASS
NM/NEOC, ppmvC	0.5		2.03	2.12	4.78	PASS

			Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL					
CO, ppmvC	0.4		1036	1045	0.85	PASS
CH ₄ , ppmvC	0.3		1068	1041	-2.49	PASS
CO ₂ , ppmvC	0.4		1022	1035	1.29	PASS
C ₂ H ₄ , ppmvC	0.4		NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4		1044	1030	-1.34	PASS
NM/NEOC, ppmvC	0.5		1024	1004	-1.96	PASS

DATE ANALYZED	11/17/2016
REFERENCE NO:	16QM2AD
	QM2-101-131



**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DIS'
 INV
 LAF
LABOI

WO #: 1631317



TO: SCAQMD LAB: OTHER: _____
 SOURCE NAME: Southern California Gas Co. I.D. No. _____
 Source Address: 12801 Tampa Ave City: Porter Ranch
 Mailing Address: _____ City: _____ Zip: 91326
 Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 11/09/16
 Approved by: Jason Low Office: _____ Budget #: 44716
 REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill
 Suspected Violation Rule(s) _____ Other _____

Sample Collected by: Kayla Jordan Date: 11/09/16 Time: 02:35pm

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Reseda Station	54214	11/08/16 / 00:00 / 24 hours	-30"	10"

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Kayla Jordan</i>	<i>[Signature]</i>	SCAQMD Lab	11/10/2016	8:11

Remarks: 1:3 scheduled samples from station at Reseda.
 Reseda Station – 18328 Gault St, Los Angeles, CA 91335
 GPS (34.199225, -118.532743). Right sampler sn 6113